

Temp. 53,000 uS standard  
(deg. C) cat. # 27143-49  
Conductivity (uS)

1	31,846
2	32,530
3	33,227
4	33,940
5	34,668
6	35,412
7	36,171
8	36,947
9	37,740
10	38,549
11	39,376
12	40,221
13	41,084
14	41,965
15	42,865
16	43,784
17	44,724
18	45,683
19	46,663
20	47,664
21	48,686
22	49,731
23	50,797
24	51,887
<b>25</b>	<b>53,000</b>
26	54,113
27	55,249
28	56,410
29	57,594
30	58,804
31	60,039
32	61,299
33	62,587
34	63,901
35	65,243
36	66,613
37	68,012
38	69,440
39	70,898
40	72,387
41	73,907
42	75,459
43	77,044
44	78,662
45	80,314
46	82,000
47	83,723
48	85,481
49	87,276

Formula is -2.1% for every degree below 25 and +2.1 % for every de

Due to temperature measurement and only two digits in the correctio  
the conductivities are only precise to the the nearest thousand (1st 2



gree above 25

n formula,  
digits).